

# Communication protocol of programmable regulated power supply

**This series of instruments adopts RS232 or USB or RS485 transmission standard to communicate with the computer. The details are as follows:**

Baud rate: 1200, 2400, 4800, 9600, 19200 Default Value :9600

Start bit: 1

Data bits: 8

Check bit: none

Stop bit: 1

## Frame structure

Message sending should start with a pause interval of at least 3.5 characters; The whole message frame must be a continuous data transmission stream. If there is a pause time of more than 3.5 characters before the frame is completed, the receiving device will refresh the incomplete message and assume that the next byte is the address field of a new message. Similarly, if a new message starts with the previous message in less than 3.5 characters, the receiving device will consider it as the continuation of the previous message.

The standard structure of a frame of information is as follows:

Start	1	2	3	4	5	6	7	8	9	10	11	end
<	0	2	0	0	4	5	8	0	0	0	0	>
Start bit	Address bit	Function bit	Data bits 1			Data bits 2			Device address			End bit

## Function

1. Set voltage
2. Read voltage
3. Set current
4. Read current

Read voltage return data format: <12012000000> return 12V

Read current return data format: <14009300000> < The first digit in the back represents the CV CC state, 1 is the CV state, C is the CC state, and the following value is the current value 9.3A

<12000000000> The return voltage is 0 CV

<C4000000000> The return current is 0 CC

ASCII send <01012100100> , Indicates that the set voltage is 12.100100V

Send in hexadecimal 3C 30 31 30 31 32 31 30 30 31 30 30 3E

Here 30 means the default is the broadcast address, and 31 means the data returned by MCU

The data format is fixed in length, and the insufficient length is supplemented by 0

PC terminal reading voltage

example: <12004580000>

Indicates that the voltage on the reading dial is 4.58V

PC terminal reading current

example: <14000183000>

Indicates that the current on the reading dial is 0.183A

PC connection (sent by PC)

<09100000000>

<01004580000>

<03006920000>

PC disconnected (PC sent to MCU)

<09200000000>

PC query voltage and current

<04003300000> Query current MCU return <14000183000>

<02012200000> Query voltage MCU return <12004580000>

PC set voltage

<01004580000> Set voltage MCU return<11OK0000000>

PC set current

<03006920000> Set current MCU return <13OK0000000>

PC startup power <07000000000>

PC power off <08000000000>